

LABORATOIRE DE
CONTROLE QUALITE

Responsable
Manager

J.FOSSE

Techniciens
Technicians

M. DUFOUR

L. FORTRIE

S. MONSAURET

T. LEURIDAN

A.MOREIRA

C.CARON

H. DEPOORTER

S.BAILLEUL

CERTIFICAT D'ANALYSE
ANALYSIS CERTIFICATE

PRODUIT : *Bandelettes de contrôle* **LOT** : 202006
pour Stéranios (2%-2%NG-2%ECS) BATCH NUMBER

PRODUCT *Stéranios control strips*

Date d'analyse : 14/02/2018
Analysis date

Vérification de l'efficacité des bandelettes à l'aide de solutions de Glutaraldéhyde ajustées à 1,5% et à 2,0%.
Verification of strips' efficacy using 1,5% and 2,0% Glutaraldéhyde solutions.

Contrôle du virage à l'orange avec bordure rose (1,5%) en 15 secondes effectué sur 10 bandelettes.

Verification of negative answer at 1,5% (Orange with pink border colouring within 15 seconds) performed with 10 strips.

Contrôle du virage au rose uniforme (2,0%) en 15 secondes effectué sur 10 bandelettes.

Verification of positive answer at 2,0% (Uniform pink colouring within 15 seconds) performed with 10 strips.


PRODUIT CONFORME
PRODUCT CONFORMITY

OUI
YES

NON
NO

Date d'émission
Sending out date
18/04/2018

Visa Technicien Qualité
Quality Technician Visa



Visa Responsable Contrôle Qualité
Quality Control Manager Visa

