



Abbott Diabetes Care Limited
Rengo Road
Witney
Oxon OX29 0YL

Tel: 0870 400 2600
Fax: 0870 400 2800

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY : G3CP
 PRODUCT NAME : FREESTYLE OPTIUM BLOOD GLUCOSE TEST STRIPS
 PACK SIZE : 50
 RE-ORDER / CATALOGUE NUMBER : 71498-75
 LOT NUMBER : 4500189273
 MANUFACTURE DATE : 2017/07/26
 EXPIRY DATE : 2018/12/31
 NUMBER OF PACKS FOR RELEASE : 10524

PRODUCT TEST DATA

CALIBRATION DATA
 INTERCEPT: 0.133 μ C SLOPE: 0.00629 μ C/mg/dL
 0.11328 μ C/mmol/L

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANGES:		TARGET:	
HIGH CONTROL	: 216 - 368	291	mg/dL
	12.0 - 20.3	16.1	mmol/L
LOW CONTROL	: 31 - 61	46	mg/dL
	1.7 - 3.4	2.6	mmol/L
MID CONTROL	: 68 - 124	96	mg/dL
	3.8 - 6.9	5.3	mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test, and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Jonathan Belcher

DATE : 24-Aug-2017 13:39

C. Lapa 24/08/17
 Caroline Deamo Lapa
 Farmaceutica
 CRF-SP 57.010

Registered Office in UK:
 Abbott Diabetes Care Limited
 Rengo Road, Witney,
 Oxon, OX29 0YL
 Registered No. 2543046
 Vat Reg No. GB 537 1507 52

