



Abbott Diabetes Care Limited  
 Range Road  
 Witney  
 Oxon OX29 0YL

Tel: 0870 400 2600  
 Fax: 0870 400 2600

**CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE**

FAMILY : G3C15  
 PRODUCT NAME : FREESTYLE OPTIUM BLOOD GLUCOSE TEST STRIPS  
 .  
 PACK SIZE : 25  
 RE-ORDER / CATALOGUE NUMBER : 71497-75  
 LOT NUMBER : 4500169729  
 MANUFACTURE DATE : 2017/11/10  
 EXPIRY DATE : 2019/04/30  
 NUMBER OF PACKS FOR RELEASE : 3231

**PRODUCT TEST DATA**

**CALIBRATION DATA**

INTERCEPT: 0.133  $\mu\text{C}$  SLOPE: 0.00629  $\mu\text{C}/\text{mg/dL}$   
 0.11328  $\mu\text{C}/\text{mmol/L}$

**EXPECTED RESULTS FOR USE WITH CONTROLS**

ASSIGNED RANGES:	TARGET:
HIGH CONTROL : 216 - 366	291 mg/dL
12.0 - 20.3	16.1 mmol/L
LOW CONTROL : 31 - 61	46 mg/dL
1.7 - 3.4	2.5 mmol/L
MID CONTROL : 68 - 124	96 mg/dL
3.8 - 6.9	5.3 mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test, and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Paul Barnes DATE : 14-Dec-2017 9:10

*C. Lapa 0510/11/18*  
 Caroline Deamo Lapa  
 Farmacêutica  
 CRF-SP 57.010

Registered Office in UK:  
 Abbott Diabetes Care Limited  
 Range Road, Witney,  
 Oxon OX29 0YL  
 Registered No. 2510046  
 Vat Reg No. GB 597 1507 32

