



Abbott Diabetes Care Limited
 Range Road
 Witney
 Oxon OX29 0YL

Tel: 0970 400 2600
 Fax: 0970 400 2600

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY : G3CHP
 PRODUCT NAME : FREESTYLE OPTIUM NEO H BLOOD GLUCOSE TEST STRIPS
 PACK SIZE : 100
 RE-ORDER / CATALOGUE NUMBER : 71312-75
 LOT NUMBER : 4500169778
 MANUFACTURE DATE : 2017/11/21
 EXPIRY DATE : 2019/04/30
 NUMBER OF PACKS FOR RELEASE : 11437

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT: 0.133 μ C SLOPE: 0.00650 μ C/mg/dL
 0.11719 μ C/mmol/L

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANGES:		TARGET:	
HIGH CONTROL	: 238 - 387	313	mg/dL
	13.2 - 21.5	17.4	mmol/L
LOW CONTROL	: 31 - 61	46	mg/dL
	1.7 - 3.4	2.6	mmol/L
MID CONTROL	: 68 - 124	96	mg/dL
	3.8 - 6.9	5.4	mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test, and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Mark Brown

DATE : 20-Dec-2017 12:23

C. Lapa 02/10/18
 Caroline Deamo Lapa
 Farmacêutica
 CRF-SP 57.010

Registered Office in UK:
 Abbott Diabetes Care Limited
 Range Road, Witney,
 Oxon OX29 0YL
 Registered No: 2548048
 Vat Reg No: GB 537 1507 52

