



Abbott Diabetes Care Limited
 Range Road
 Witney
 Oxon OX29 0YL

Tel: 0870 400 2600
 Fax: 0870 400 2800

CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE

FAMILY : G3CH
PRODUCT NAME : FREESTYLE OPTIUM H BLOOD GLUCOSE TEST STRIPS

PACK SIZE : 50
RE-ORDER / CATALOGUE NUMBER : 80127-70
LOT NUMBER : 4500169799
MANUFACTURE DATE : 2017/11/24
EXPIRY DATE : 2019/04/30
NUMBER OF PACKS FOR RELEASE : 23255

PRODUCT TEST DATA

CALIBRATION DATA

INTERCEPT:	0.359	μC	SLOPE:	0.03447	$\mu\text{C}/\text{mg}/\text{dL}$
				0.62109	$\mu\text{C}/\text{mmol}/\text{L}$

EXPECTED RESULTS FOR USE WITH CONTROLS

ASSIGNED RANGES:				TARGET:	
HIGH CONTROL	:	227	-	377	302 mg/dL
		12.6	-	20.9	16.8 mmol/L
LOW CONTROL	:	32	-	63	48 mg/dL
		1.8	-	3.5	2.7 mmol/L
MID CONTROL	:	67	-	123	95 mg/dL
		3.7	-	6.8	5.3 mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test, and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Paul Barnes **DATE :** 05-Jan-2018 15:50

Registered Office in UK:
 Abbott Diabetes Care Limited
 Range Road, Witney,
 Oxon OX29 0YL
 Registered No. 2548046
 Vat Reg No. GB 537 1507 52

