



Abbott Diabetes Care Limited  
 Range Road  
 Witney  
 Oxon OX29 9YL

Tel: 0870 400 2890  
 Fax: 0570 400 2890

**CERTIFICATE OF CONFORMITY AND APPROVAL FOR RELEASE**

FAMILY : G3CHP  
 PRODUCT NAME : FREESTYLE OPTIUM NEO H BLOOD GLUCOSE TEST STRIPS

PACK SIZE : 100  
 RE-ORDER / CATALOGUE NUMBER : 71312-75  
 LOT NUMBER : 4500169819  
 MANUFACTURE DATE : 2017/11/29  
 EXPIRY DATE : 2019/04/30  
 NUMBER OF PACKS FOR RELEASE : 11331

<u>CALIBRATION DATA</u>		<u>PRODUCT TEST DATA</u>			
INTERCEPT:	0.133	µC	SLOPE:	0.00645	µC/mg/dL
				0.11621	µC/mmol/L

**EXPECTED RESULTS FOR USE WITH CONTROLS**

<u>ASSIGNED RANGES:</u>		<u>TARGET:</u>				
HIGH CONTROL	:	234	-	384	309	mg/dL
		13.0	-	21.3	17.2	mmol/L
LOW CONTROL	:	31	-	61	46	mg/dL
		1.7	-	3.4	2.6	mmol/L
MID CONTROL	:	68	-	124	96	mg/dL
		3.8	-	6.9	5.4	mmol/L

The above product has been manufactured in accordance with Abbott Diabetes Care Operating Procedures and conforms in all respects with the specific finished goods specifications as tested and inspected by Product Test, and conforms to order part numbers as defined by the Quality System.

AUTHORISED BY : Mark Brown

DATE : 03-Jan-2018 14:10

*[Signature]*  
 Danilo Coimbra Silverio  
 Farmacêutico  
 CRF-SP 41.734

Registered Office in UK:  
 Abbott Diabetes Care Limited  
 Range Road, Witney,  
 Oxon OX29 9YL  
 Registered No. 2549046  
 Vit Reg No. G5 537 1607 52

